



CSA INTERNATIONAL

Certificate of Compliance

Certificate: 1868261

Master Contract: 237868

Project: 2007815

Date Issued: 2008/05/16

Issued to: Korea New Ceramics Co. Ltd.

70-3 Sosabon-2 dong

Sosa-Gu

Bucheon, 422-807

South Korea

Attention: Mr. Kinam Kim

The products listed below are eligible to bear the CSA Mark shown with adjacent indicator ▲



Issued by: A. Holbeche P. Eng

Authorized by: Patricia Pasemko, Operations Manager

PRODUCTS

CLASS 4828 31 - SIGNAL APPLIANCES-Combustible Gas Detection Instruments - -For Hazardous Locations-Components

LEL Sensor, Model Nos. KGS 701, 702 and 703 and Models KGS 801, 802 and 803, plug-in, open terminal, gas-sensors with a sintered metal disk intended for assembly as part of an intrinsically safe portable combustible gas detector.

Note: The hazardous locations rating is determined by the rating of the equipment it is installed in.



Certificate: 1868261

Master Contract: 237868

Project: 2007815

Date Issued: 2008/05/16

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 0-M91 - General Requirements-Canadian Electrical Code, Part II

CSA Std C22.2 No. 30-M1986 - Explosion-Proof Enclosures for Use in Class I Hazardous Locations

MARKINGS

Appear on a CSA approved label system:

- Submitter's identification.
- Model designation or equivalent identification.
- Date code or serial number.
- CSA Component Acceptance marking



Supplement to Certificate of Compliance

Certificate: 1868261

Master Contract: 237868

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
2007815	2008/05/16	Update to Report 1868261 for additional models KGS 801, 802, 803. Alternate epoxy resin added.